

Rendelési információk:

Katalógusszám Kiszerezés
50911 1 x 5 ml

Felhasználás:

Kreatinin reagens kalibrálására szolgáló stabil folyékony kreatinin standard.
Felhasználásra kész.
A standard koncentrációja LOT specifikus.

Visszavezethetőség:

GC-IDMS

Stabilitás:

Felbontás nélkül 2-8°C-on tárolva a lejáratí időig.
Felbontás után 2-8°C-on tárolva minimum 4 hétig.

Figyelmeztetések:

Csak in vitro diagnosztikai célokra használható.
Ne használja a standardet a lejáratí dátum eltelte után!

A címkéken a következő szimbólumok lehetnek

	In vitro diagnosztikai orvostechnikai eszköz
	Gyártó
	CE-jelölés
	Tárolási hőmérséklet
	Lejáratí idő (év/hónap)
	Gyártási szám
	Kódszám
	Tárolási irány
	Biológiai veszély

Ordering information:

Catalogue number Size
50911, 950911 1 x 5 ml

Preparation:

Creatinine standard for the calibration of creatinine reagent.
Ready to use.
Concentration is LOT specific.

Traceability:

GC-IDMS










Stability:

Without opening at 2-8°C until the expiry date.
After first opening at least 4 weeks at 2-8°C.

Warnings:

For in vitro diagnostic use only.
Do not use after the expiry date!

The following symbols can be used on the labels

	In vitro diagnostic device
	Manufacturer
	CE-marking
	Temperature limitations
	Use by (year/month)
	Batch code
	Catalogue number
	This way up
	Biological risk

Verzió/Version: 17-2015-06

Felülvizsgálat dátuma/Date of revision : 2016-03