

**Felhasználás:**

Összefehérje (vizelet) reagens kalibrálására szolgáló stabil folyékony vizeletfehérje standard.

Felhasználásra kész.

A standard koncentrációja LOT specifikus.

**Visszavezethetőség:**

SRM 927c

**Stabilitás:**

Felbontás nélkül 2-8°C-on tárolva a lejárati időig.

Felbontás után 2-8°C-on tárolva minimum 4 hétig.

**Figyelmeztetések:**

Csak in vitro diagnosztikai célokra használható.

Ne használja a standardet a lejárati dátum eltelte után!

**A címkeken a következő szimbólumok lehetnek**

	In vitro diagnosztikai orvostechnikai eszköz
	Gyártó
	CE-jelölés
	Tárolási hőmérséklet
	Lejárati idő (év/hónap)
	Gyártási szám
	Kódszám
	Tárolási irány
	Biológiai veszély

**Preparation:**

Total protein ultrasensitive standard for the calibration of total protein ultrasensitive reagent.

Ready to use.

Concentration is LOT specific.

**Traceability:**

SRM 927c

**Stability:**

Without opening at 2-8°C until the expiry date.










After first opening at least 4 weeks at 2-8°C.

**Warnings:**

For in vitro diagnostic use only.

Do not use after the expiry date!

**The following symbols can be used on the labels**

	In vitro diagnostic device
	Manufacturer
	CE-marking
	Temperature limitations
	Use by (year/month)
	Batch code
	Catalogue number
	This way up
	Biological risk

Verzió/Version: 32B-2015-06

Felülvizsgálat dátuma/Date of revision : 2016-03