


**Reagent kit for the quantitative determination of total protein concentration in serum. Biuret method.**

The human body contains countless different protein (50% in cells). Not only is the variety of proteins seemingly infinite, so are their variations of concentration in health and disease, their distribution within the body, their functions, their compositions and their structures. Most plasma proteins with the exception of immunoglobulins and hormonal proteins are synthesized in liver. They function as major components of cells, are involved in transport, enzyme catalysis, homeostatic control, hormonal regulation, blood coagulation, immunity, growth and repair, and heredity.

**Principle**

Cupric ions in an alkaline solution react with the peptide bonds of proteins and polypeptides containing at least two peptide bonds to produce a violet colored complex. The absorbance of the complex at 546 nm is directly proportional to the concentration of protein in the sample.



**Reference values**

**Serum total protein** 62-80 g/l (6,2-8,0 g/dl)

It is recommended that each laboratory should assign its own normal range.

**Reagents**

**1.Reagent(R1)**

Potassium iodide	30 mmol/l
Potassium sodium tartrate	100 mmol/l
Copper sulphate	30 mmol/l
Sodium hydroxide	3.8 mol/l

**Safety instructions:**

**Reagent 1:**

X, Irritative

R36/38 Irritating to eyes and skin

S26 In case of contact with eyes, rinse immediately with plenty of water and seek medical advice

**Sample**

Serum free of haemolysis.

**Note**

This reagent contains sodium azide. To avoid the possible build-up of azide compounds, flush waste-pipes with water after the disposal of undiluted reagent.

For greater accuracy a serum blank correction is recommended when assaying turbid or lipaemic samples.

**PROCEDURE**

**Preparation and stability of working reagent**

The reagents are ready for use.

If the absorbance of working reagent is higher than 0.1 at 546 nm the reagent can not be used.

**Assay conditions**

Wavelength:	546 nm (530-580) nm
Temperature:	37 °C
Cuvette:	1 cm light path
Read against:	reagent blank
Method:	endpoint (increasing)

**Pipette into cuvette**

	Blank	Standard	Sample
<b>Reagent</b>	1 ml	1 ml	1 ml
<b>Distilled water</b>	10µl		
<b>Standard</b>		10µl	
<b>Sample</b>			10µl

Mix and read the absorbance (A) after a ten-minute incubation.

**Calibration (37°C, Biuret method)**

S1: Distilled water

S2: Diagnosticum DunaCal Cat. No.: Dcal or

Total protein standard Cat. No.: 951911 or

Roche C.F.A.S. (Calibrator for automated system)

Randox Calibration Serum Level I or

Randox Calibration Serum Level II

**Calibration frequency**

Two-point calibration is recommended:

- after reagent lot change,

- as required following quality control procedures.

**Calculation using calibration**

# TOTAL PROTEIN (BIURET)

## STABLE LIQUID REAGENT

CE

Cat. No.:	911953	911955	911993
	1x250 ml	20x20 ml	2x500 ml

$$\frac{A_{sample}}{A_{standard}} \times C_{standard} = C_{sample}$$

A = Absorbance, C = Concentration

**Quality control**

A quality control program is recommended for all clinical laboratories. The analysis of control material in both the normal and abnormal ranges with each assay is recommended for monitoring the performance of the procedure. Each laboratory should establish corrective measures to be taken if values fall outside the limits. Recommended controls: Diagnosticum DunaCont N (Cat. No.: Dcon-N) and DunaCont P (Cat. No.: Dcon-P)

**PERFORMANCES DATA**

The following data were obtained using the Olympus 600 analyzers.

**Linearity**

The test is linear up to 120g/l (12,0 g/dl) protein concentration.

**Sensitivity**

It is recommended that each laboratory establishes its own range of sensitivity as this is limited by the sensitivity of the spectrophotometer used. Under manual conditions however, a change of 0.001 Abs is equivalent to 0.18 g/l (0,018 g/dl) Protein concentration at 546 nm.

**Precision**

Sample	Reproducibility		
	Average conc. (g/l)	SD	CV%
Sample I.	50.1	0.462	0.92
Sample II.	51.8	0.172	0.33

**Correlation**

Comparative studies were done to compare our reagent with another commercial Total Protein reagent on 60 human samples.

The results from these studies are detailed below.

Correlation coefficient:  $r=0.9957$

Linear regression:  $y \text{ (g/l)} = 0.994x + 0.525$

( $x$  = other commercial reagent,  $y$  = own reagent).

**Specificity**

Bilirubin 855 µmol/l (50mg/dl), lipid 1000mg/dl, glucose 55.5 mmol/l (1000mg/dl) and ascorbic acid 2.84 mmol/l (50mg/dl) don't interfere with the assay up to the given levels.

**NOTE**

Do not use reagents after the expiry date stated on each reagent container label. Do not use products, test solutions and reagents described above for any purpose other than described herein.

**For *in vitro* diagnostic use only.**

**The following symbols are used on labels**

 For *in vitro* diagnostic use

 Use by (last day of the month)

 Temperature limitation

 Batch Code

 Code

**Bibliography**

Gornall A. et al.: *J. Biol. Chem.* 177, 751 (1949)

Weichselbaum, P.E.: *Am. J. Path.* 16, 40 (1946)